

# Certificate of Compliance

Certificate: 1968359

Master Contract: 219283

Project: 2332133

Date Issued: December 1, 2010

Issued to: Kongsberg Maritime AS  
Haakon VII's gate 4  
N-7005 Trondheim  
NORWAY

Attention: Mr. Jan Stewart Rambech

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only*



V. Andersen

Issued by: V. Andersen

## PRODUCTS

CLASS 2252 05 - PROCESS CONTROL EQUIPMENT

CLASS 2252 85 - PROCESS CONTROL EQUIPMENT (Certified to U.S. Standards)

Kongsberg wireless temperature monitoring system, model GB... 733 and model GB... 110

Permanently connected, rated 24VDC, 0.33A. Relay load: Max 1A, 125VAC.

where each "." can be A-Z and/or 0-9

The differences are only software related.

**Certificate:** 1968359

**Project:** 2332133



**Master Contract:** 219283

**Date Issued:** December 1, 2010

---

### **APPLICABLE REQUIREMENTS**

CAN/CSA-C22.2 No. 61010.1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

UL Std. No. 61010-1 (2<sup>nd</sup> Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

### **MARKINGS**

The following markings appear on the product and may be described in further detail in the Certification Report Supplement, Section C:

1. Submitter's identification (company name and/or file number and/or registered tradename);
2. Model designation;
3. Electrical rating;
4. Date of manufacture (or traceable serial number);
5. The CSA Monogram with C/US subscript;
6. The ISO 7000 – 0434 caution symbol;



CSA INTERNATIONAL

## *Supplement to Certificate of Compliance*

**Certificate:** 1968359

**Master Contract:** 219283

*The products listed, including the latest revision described below,  
are eligible to be marked in accordance with the referenced Certificate.*

### **Product Certification History**

---

<b>Project</b>	<b>Date</b>	<b>Description</b>
2332133	December 1, 2010	Update to certificate 1968359 considers alternative components.
1968359	December 21, 2007	Original cCSAus Certification of Kongsberg wireless temperature monitoring system, model GB... 733 and model GB... 110 where each "." can be A-Z and/or 0-9. The differences are only software related.